

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
 Data File : VV032112.D
 Acq On : 15 Sep 2023 11:32
 Operator : SY/MD
 Sample : VV0915MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK394

Quant Time: Sep 16 02:11:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 16 02:09:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	215977	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	207107	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	105986	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	48446	35.842	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.680%
7) Chloroethane-d5	1.523	69	46429	39.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.360%
11) 1,1-Dichloroethene-d2	2.057	65	22148	34.994	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.980%
21) 2-Butanone-d5	3.783	46	115373	108.527	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.530%
24) Chloroform-d	4.246	84	124641	42.189	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.380%
26) 1,2-Dichloroethane-d4	4.941	65	79934	40.770	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.540%
32) Benzene-d6	4.960	84	237805	42.079	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.160%
36) 1,2-Dichloropropane-d6	5.989	67	86550	47.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.160%
41) Toluene-d8	7.243	98	206672	39.690	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.380%#
43) trans-1,3-Dichloroprop...	7.555	79	38030	40.338	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.680%
47) 2-Hexanone-d5	8.027	63	69788	125.178	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	125.180%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	123802	52.669	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.340%
66) 1,2-Dichlorobenzene-d4	11.564	152	90177	46.674	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.340%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
Data File : VV032112.D
Acq On : 15 Sep 2023 11:32
Operator : SY/MD
Sample : VV0915MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK394

Quant Time: Sep 16 02:11:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 16 02:09:37 2023
Response via : Initial Calibration

