

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091523\
 Data File : VW032124.D
 Acq On : 15 Sep 2023 16:12
 Operator : SY/MD
 Sample : PB155573TB 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VLEB573

Quant Time: Sep 16 02:14:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 16 02:09:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	209121	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	205887	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	106806	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	77751	59.408	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	118.820%
7) Chloroethane-d5	1.532	69	62765	54.708	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.420%
11) 1,1-Dichloroethene-d2	2.059	65	31635	51.622	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.240%
21) 2-Butanone-d5	3.793	46	105375	102.372	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.370%
24) Chloroform-d	4.256	84	130853	45.743	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.480%
26) 1,2-Dichloroethane-d4	4.947	65	83239	43.847	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.700%
32) Benzene-d6	4.963	84	277685	49.427	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.860%
36) 1,2-Dichloropropane-d6	5.992	67	90107	49.831	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.660%
41) Toluene-d8	7.246	98	251540	48.593	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.180%
43) trans-1,3-Dichloroprop...	7.558	79	39379	42.016	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.040%
47) 2-Hexanone-d5	8.027	63	67750	122.242	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	122.240%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	116786	49.979	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.960%
66) 1,2-Dichlorobenzene-d4	11.564	152	94771	48.675	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
Data File : VV032124.D
Acq On : 15 Sep 2023 16:12
Operator : SY/MD
Sample : PB155573TB 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VLEB573

Quant Time: Sep 16 02:14:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 16 02:09:37 2023
Response via : Initial Calibration

