

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091523\
 Data File : VW032115.D
 Acq On : 15 Sep 2023 12:45
 Operator : SY/MD
 Sample : 04359-01 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JYCA2

Quant Time: Sep 16 02:12:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 16 02:09:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	212661	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	207496	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	105676	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	87844	66.003	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 132.000%	
7) Chloroethane-d5	1.532	69	70006	60.004	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 120.000%	
11) 1,1-Dichloroethene-d2	2.060	65	35166	56.428	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 112.860%	
21) 2-Butanone-d5	3.789	46	111610	106.625	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 106.620%	
24) Chloroform-d	4.252	84	146779	50.457	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 100.920%	
26) 1,2-Dichloroethane-d4	4.944	65	93681	48.526	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 97.060%	
32) Benzene-d6	4.963	84	309722	54.702	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.400%	
36) 1,2-Dichloropropane-d6	5.992	67	103928	57.028	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 114.060%	
41) Toluene-d8	7.246	98	277165	53.128	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 106.260%	
43) trans-1,3-Dichloroprop...	7.555	79	43519	46.073	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 92.140%	
47) 2-Hexanone-d5	8.027	63	72968	130.636	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 130.640%#	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	127895	54.309	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 108.620%	
66) 1,2-Dichlorobenzene-d4	11.564	152	103377	53.663	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 107.320%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
Data File : VV032115.D
Acq On : 15 Sep 2023 12:45
Operator : SY/MD
Sample : 04359-01 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JYCA2

Quant Time: Sep 16 02:12:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 16 02:09:37 2023
Response via : Initial Calibration

