

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091523\
 Data File : VW032119.D
 Acq On : 15 Sep 2023 14:15
 Operator : SY/MD
 Sample : 04359-07 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLAD5

Quant Time: Sep 16 02:13:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 16 02:09:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	218752	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	209139	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	107816	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	80527	58.821	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	117.640%
7) Chloroethane-d5	1.532	69	65075	54.224	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.440%
11) 1,1-Dichloroethene-d2	2.060	65	33432	52.152	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.300%
21) 2-Butanone-d5	3.793	46	106442	98.856	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.860%
24) Chloroform-d	4.256	84	136363	45.571	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.140%
26) 1,2-Dichloroethane-d4	4.947	65	87377	44.001	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.000%
32) Benzene-d6	4.963	84	289733	50.770	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.540%
36) 1,2-Dichloropropane-d6	5.992	67	96630	52.607	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.220%
41) Toluene-d8	7.246	98	261328	49.699	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.400%
43) trans-1,3-Dichloroprop...	7.558	79	40369	42.403	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.800%
47) 2-Hexanone-d5	8.027	63	69163	122.851	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	122.850%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	119925	50.524	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.040%
66) 1,2-Dichlorobenzene-d4	11.564	152	98614	50.174	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
Data File : VV032119.D
Acq On : 15 Sep 2023 14:15
Operator : SY/MD
Sample : 04359-07 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLAD5

Quant Time: Sep 16 02:13:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 16 02:09:37 2023
Response via : Initial Calibration

