

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
 Data File : VV032120.D
 Acq On : 15 Sep 2023 14:38
 Operator : SY/MD
 Sample : 04359-05 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JLAD3

Quant Time: Sep 16 02:13:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 16 02:09:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	197314	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	190344	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	98910	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	76668	62.086	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	124.180%
7) Chloroethane-d5	1.532	69	61502	56.815	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.620%
11) 1,1-Dichloroethene-d2	2.060	65	30917	53.469	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.940%
21) 2-Butanone-d5	3.789	46	106514	109.671	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.670%
24) Chloroform-d	4.253	84	131049	48.553	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.100%
26) 1,2-Dichloroethane-d4	4.947	65	84474	47.161	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.320%
32) Benzene-d6	4.963	84	273881	52.731	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.460%
36) 1,2-Dichloropropane-d6	5.992	67	91281	54.602	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.200%
41) Toluene-d8	7.246	98	246141	51.433	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.860%
43) trans-1,3-Dichloroprop...	7.555	79	39671	45.784	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.560%
47) 2-Hexanone-d5	8.027	63	67473	131.684	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	131.680%#
56) 1,1,2,2-Tetrachloroeth...	10.156	84	118685	54.939	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.880%
66) 1,2-Dichlorobenzene-d4	11.564	152	92334	51.209	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.420%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091523\
Data File : VV032120.D
Acq On : 15 Sep 2023 14:38
Operator : SY/MD
Sample : 04359-05 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JLAD3

Quant Time: Sep 16 02:13:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082323WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 16 02:09:37 2023
Response via : Initial Calibration

