

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091619\
 Data File : VV012858.D
 Acq On : 17 Sep 2019 01:20
 Operator : SY/MD
 Sample : K4834-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Sep 17 08:50:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 17 07:22:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	385029	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	381148	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	137586	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	120941	4.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.60%
7) Chloroethane-d5	1.58	69	98416	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.13	63	157286	3.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.20%
20) 2-Butanone-d5	3.97	46	197163	53.49	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.98%
24) Chloroform-d	4.40	84	203298	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	5.08	65	101316	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	5.09	84	407777	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.12	67	126747	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.36	98	314422	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	7.66	79	36511	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
46) 2-Hexanone-d5	8.14	63	112273	48.92	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.84%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	72465	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.00%
64) 1,2-Dichlorobenzene-d4	11.67	152	105269	5.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.40%

Target Compounds					Qvalue
13) Acetone	2.23	43	16678m	5.840	ug/L

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091619\
Data File : VV012858.D
Acq On : 17 Sep 2019 01:20
Operator : SY/MD
Sample : K4834-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Sep 17 08:50:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 17 07:22:21 2019
Response via : Initial Calibration

