

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091619\
 Data File : VV012834.D
 Acq On : 16 Sep 2019 14:06
 Operator : SY/MD
 Sample : K4780-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BM3

Quant Time: Sep 16 17:19:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 17:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	391354	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	393441	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	148311	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	206441	7.62	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	152.40%#
7) Chloroethane-d5	1.58	69	169425	7.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	159.40%#
11) 1,1-Dichloroethene-d2	2.13	63	256023	5.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.00%
20) 2-Butanone-d5	3.98	46	450473	120.24	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	240.48%#
24) Chloroform-d	4.40	84	324995	7.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	152.60%#
26) 1,2-Dichloroethane-d4	5.08	65	171626	8.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	175.20%#
32) Benzene-d6	5.09	84	674691	7.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	157.00%#
36) 1,2-Dichloropropane-d6	6.12	67	210300	8.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	160.80%#
41) Toluene-d8	7.36	98	540980	7.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	140.20%#
43) trans-1,3-Dichloropropene-	7.66	79	58550	6.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	138.60%#
46) 2-Hexanone-d5	8.14	63	225947	95.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	190.74%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	117342	8.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	163.20%#
64) 1,2-Dichlorobenzene-d4	11.67	152	186719	9.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	181.60%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091619\
Data File : VV012834.D
Acq On : 16 Sep 2019 14:06
Operator : SY/MD
Sample : K4780-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BM3

Quant Time: Sep 16 17:19:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 17:02:53 2019
Response via : Initial Calibration

