

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091619\
 Data File : VV012835.D
 Acq On : 16 Sep 2019 14:33
 Operator : SY/MD
 Sample : K4780-12
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BN7

Quant Time: Sep 16 17:23:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 16 17:02:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	394530	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	396678	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	147587	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	209828	7.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	153.60%#
7) Chloroethane-d5	1.58	69	171123	7.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	159.80%#
11) 1,1-Dichloroethene-d2	2.13	63	254472	5.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	112.40%
20) 2-Butanone-d5	3.98	46	443767	117.50	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	235.00%#
24) Chloroform-d	4.40	84	325326	7.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	151.60%#
26) 1,2-Dichloroethane-d4	5.08	65	172026	8.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	174.20%#
32) Benzene-d6	5.09	84	670176	7.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	154.60%#
36) 1,2-Dichloropropane-d6	6.12	67	209812	7.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	159.00%#
41) Toluene-d8	7.36	98	537967	6.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	138.20%#
43) trans-1,3-Dichloropropene-	7.66	79	58962	6.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	138.40%#
46) 2-Hexanone-d5	8.14	63	213079	89.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	178.40%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	120397	8.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	166.00%#
64) 1,2-Dichlorobenzene-d4	11.67	152	187688	9.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	183.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091619\
Data File : VV012835.D
Acq On : 16 Sep 2019 14:33
Operator : SY/MD
Sample : K4780-12
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BN7

Quant Time: Sep 16 17:23:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 16 17:02:53 2019
Response via : Initial Calibration

