

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091619\
 Data File : VV012841.D
 Acq On : 16 Sep 2019 17:23
 Operator : SY/MD
 Sample : K4780-11RE
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BM3RE

Quant Time: Sep 17 07:33:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 17 07:22:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	394207	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	396144	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	151947	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	167451	6.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.60%
7) Chloroethane-d5	1.58	69	150120	7.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	140.20%#
11) 1,1-Dichloroethene-d2	2.13	63	213787	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.60%
20) 2-Butanone-d5	3.97	46	435352	115.36	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	230.72%#
24) Chloroform-d	4.40	84	306295	7.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	142.80%#
26) 1,2-Dichloroethane-d4	5.08	65	160515	8.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	162.60%#
32) Benzene-d6	5.09	84	613738	7.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	141.80%#
36) 1,2-Dichloropropane-d6	6.12	67	195534	7.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	148.40%#
41) Toluene-d8	7.36	98	483655	6.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	124.40%
43) trans-1,3-Dichloropropene-	7.66	79	62755	7.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	147.60%#
46) 2-Hexanone-d5	8.14	63	215812	90.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	180.94%#
57) 1,1,2,2-Tetrachloroethane-	10.26	84	114713	7.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	158.40%#
64) 1,2-Dichlorobenzene-d4	11.67	152	175444	8.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	166.60%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091619\
Data File : VV012841.D
Acq On : 16 Sep 2019 17:23
Operator : SY/MD
Sample : K4780-11RE
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BM3RE

Quant Time: Sep 17 07:33:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 17 07:22:21 2019
Response via : Initial Calibration

