

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091619\  
 Data File : VV012848.D  
 Acq On : 16 Sep 2019 20:55  
 Operator : SY/MD  
 Sample : K4834-02  
 Misc : 25.0mL/MSVOA V/WATER  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 C0MD6

Quant Time: Sep 17 08:04:59 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR091619WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Sep 17 07:22:21 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.66  | 114  | 394940   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.89  | 117  | 390766   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.29 | 152  | 142899   | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.32   | 65    | 125379   | 4.58     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 91.60%  |
| 7) Chloroethane-d5             | 1.58   | 69    | 102090   | 4.76     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 95.20%  |
| 11) 1,1-Dichloroethene-d2      | 2.13   | 63    | 160156   | 3.54     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 70.80%  |
| 20) 2-Butanone-d5              | 3.97   | 46    | 203645   | 53.86    | ug/L | -0.02   |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 107.72% |
| 24) Chloroform-d               | 4.40   | 84    | 201187   | 4.68     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 93.60%  |
| 26) 1,2-Dichloroethane-d4      | 5.08   | 65    | 101970   | 5.16     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 103.20% |
| 32) Benzene-d6                 | 5.09   | 84    | 410919   | 4.81     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 96.20%  |
| 36) 1,2-Dichloropropane-d6     | 6.12   | 67    | 127018   | 4.89     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 97.80%  |
| 41) Toluene-d8                 | 7.36   | 98    | 319269   | 4.17     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 83.40%  |
| 43) trans-1,3-Dichloropropene- | 7.66   | 79    | 37217    | 4.43     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 88.60%  |
| 46) 2-Hexanone-d5              | 8.14   | 63    | 110717   | 47.05    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 94.10%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.26  | 84    | 72215    | 5.06     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 101.20% |
| 64) 1,2-Dichlorobenzene-d4     | 11.67  | 152   | 109100   | 5.51     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 110.20% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091619\  
Data File : VV012848.D  
Acq On : 16 Sep 2019 20:55  
Operator : SY/MD  
Sample : K4834-02  
Misc : 25.0mL/MSVOA V/WATER  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
C0MD6

Quant Time: Sep 17 08:04:59 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR091619WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Sep 17 07:22:21 2019  
Response via : Initial Calibration

