

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091619\
 Data File : VV012853.D
 Acq On : 16 Sep 2019 23:08
 Operator : SY/MD
 Sample : K4834-08
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COMF8

Quant Time: Sep 17 08:22:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 17 07:22:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	389417	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	385104	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	139153	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	124031	4.60	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.00%
7) Chloroethane-d5	1.58	69	102135	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.13	63	159463	3.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.40%
20) 2-Butanone-d5	3.98	46	207154	55.57	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.14%
24) Chloroform-d	4.40	84	201929	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.08	65	102738	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
32) Benzene-d6	5.09	84	411969	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	6.12	67	127748	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	7.36	98	315030	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	7.66	79	38129	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
46) 2-Hexanone-d5	8.14	63	109868	47.38	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.76%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	72710	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	107626	5.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091619\
Data File : VV012853.D
Acq On : 16 Sep 2019 23:08
Operator : SY/MD
Sample : K4834-08
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COMF8

Quant Time: Sep 17 08:22:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 17 07:22:21 2019
Response via : Initial Calibration

