

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091620\
 Data File : VV018233.D
 Acq On : 16 Sep 2020 13:32
 Operator : SY/MD
 Sample : L4012-15
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFWT3

Quant Time: Sep 17 05:01:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 17 04:38:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	200982	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	199419	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	97304	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	86709	52.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.88%
7) Chloroethane-d5	1.58	69	81113	60.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	121.04%
11) 1,1-Dichloroethene-d2	2.12	63	135306	45.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.06%
21) 2-Butanone-d5	3.91	46	140083	139.73	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	139.73%#
24) Chloroform-d	4.38	84	182146	64.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	128.20%#
26) 1,2-Dichloroethane-d4	5.06	65	120995	66.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	132.84%#
32) Benzene-d6	5.08	84	350556	61.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	122.82%
36) 1,2-Dichloropropane-d6	6.09	67	115881	63.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	127.76%#
41) Toluene-d8	7.34	98	306206	58.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.70%
43) trans-1,3-Dichloropropene-	7.64	79	56731	61.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	123.10%
47) 2-Hexanone-d5	8.11	63	103316	150.37	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	150.37%#
57) 1,1,2,2-Tetrachloroethane-	10.24	84	164178	72.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	144.22%#
64) 1,2-Dichlorobenzene-d4	11.65	152	140022	70.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	140.14%#

Target Compounds					Ovalue
13) Acetone	2.19	43	2953	3.678 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091620\
Data File : VV018233.D
Acq On : 16 Sep 2020 13:32
Operator : SY/MD
Sample : L4012-15
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFWT3

Quant Time: Sep 17 05:01:49 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 17 04:38:07 2020
Response via : Initial Calibration

