

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091620\
 Data File : VV018259.D
 Acq On : 17 Sep 2020 00:50
 Operator : SY/MD
 Sample : L4042-13
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AC4

Quant Time: Sep 17 06:14:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 17 05:33:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	291628	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	288829	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	138551	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	92017	38.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.44%
7) Chloroethane-d5	1.58	69	77329	39.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.52%
11) 1,1-Dichloroethene-d2	2.12	63	133187	30.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.10%
21) 2-Butanone-d5	3.92	46	116419	80.03	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.03%
24) Chloroform-d	4.38	84	165618	40.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.32%
26) 1,2-Dichloroethane-d4	5.06	65	110289	41.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.44%
32) Benzene-d6	5.08	84	331753	40.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.24%
36) 1,2-Dichloropropane-d6	6.09	67	105858	40.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.58%
41) Toluene-d8	7.34	98	295265	39.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.36%#
43) trans-1,3-Dichloropropene-	7.64	79	50420	37.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.54%
47) 2-Hexanone-d5	8.11	63	83892	84.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.30%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	141269	42.84	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.68%
64) 1,2-Dichlorobenzene-d4	11.65	152	125643	44.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091620\
Data File : VV018259.D
Acq On : 17 Sep 2020 00:50
Operator : SY/MD
Sample : L4042-13
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AC4

Quant Time: Sep 17 06:14:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 17 05:33:55 2020
Response via : Initial Calibration

