

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091620\
 Data File : VV018246.D
 Acq On : 16 Sep 2020 19:56
 Operator : SY/MD
 Sample : VV0916WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK71

Quant Time: Sep 17 05:40:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 17 05:33:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	291537	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	288855	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	140631	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	97368	40.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.96%
7) Chloroethane-d5	1.58	69	80508	41.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.82%
11) 1,1-Dichloroethene-d2	2.12	63	140303	32.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.38%
21) 2-Butanone-d5	3.92	46	124428	85.56	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.56%
24) Chloroform-d	4.38	84	173970	42.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.40%
26) 1,2-Dichloroethane-d4	5.06	65	115918	43.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.74%
32) Benzene-d6	5.08	84	351660	42.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.06%
36) 1,2-Dichloropropane-d6	6.10	67	113897	43.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.70%
41) Toluene-d8	7.34	98	315031	41.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.60%
43) trans-1,3-Dichloropropene-	7.64	79	53864	40.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.68%
47) 2-Hexanone-d5	8.11	63	90955	91.39	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.39%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	147164	44.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	135115	46.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091620\
Data File : VV018246.D
Acq On : 16 Sep 2020 19:56
Operator : SY/MD
Sample : VV0916WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK71

Quant Time: Sep 17 05:40:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 17 05:33:55 2020
Response via : Initial Calibration

