

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091621\
 Data File : VW022161.D
 Acq On : 16 Sep 2021 16:56
 Operator : SY/MD
 Sample : M3764-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0C97

Quant Time: Sep 17 01:57:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 17 01:54:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	300482	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	285905	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	129073	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	89663	43.089	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.180%
7) Chloroethane-d5	1.561	69	79472	48.906	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.820%
11) 1,1-Dichloroethene-d2	2.105	63	134542	39.095	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.200%
21) 2-Butanone-d5	3.876	46	155537	108.988	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.990%
24) Chloroform-d	4.352	84	189144	49.850	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.700%
26) 1,2-Dichloroethane-d4	5.034	65	119027	53.258	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.520%
32) Benzene-d6	5.050	84	392248	50.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.160%
36) 1,2-Dichloropropane-d6	6.072	67	124167	50.914	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.820%
41) Toluene-d8	7.317	98	343106	46.871	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.740%
43) trans-1,3-Dichloroprop...	7.625	79	50465	43.476	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.960%
47) 2-Hexanone-d5	8.088	63	96801	103.448	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.450%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	157097	50.297	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	130640	51.485	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091621\
Data File : VV022161.D
Acq On : 16 Sep 2021 16:56
Operator : SY/MD
Sample : M3764-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0C97

Quant Time: Sep 17 01:57:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 17 01:54:12 2021
Response via : Initial Calibration

