

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091621\
 Data File : VV022168.D
 Acq On : 16 Sep 2021 20:09
 Operator : SY/MD
 Sample : M3764-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CA8

Quant Time: Sep 17 01:58:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 17 01:54:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	307989	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	290937	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	127934	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	87987	41.253	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.500%
7) Chloroethane-d5	1.564	69	78920	47.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.760%
11) 1,1-Dichloroethene-d2	2.105	63	133443	37.831	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.660%
21) 2-Butanone-d5	3.905	46	163643	111.873	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.870%
24) Chloroform-d	4.352	84	193750	49.819	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.640%
26) 1,2-Dichloroethane-d4	5.034	65	120296	52.514	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.020%
32) Benzene-d6	5.050	84	396482	50.244	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.480%
36) 1,2-Dichloropropane-d6	6.072	67	126327	50.904	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.800%
41) Toluene-d8	7.320	98	351771	47.223	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.440%
43) trans-1,3-Dichloroprop...	7.625	79	51290	43.423	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.840%
47) 2-Hexanone-d5	8.092	63	96849	101.709	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.710%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	156080	49.107	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.220%
66) 1,2-Dichlorobenzene-d4	11.628	152	131891	52.441	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.880%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.114	96	3821	2.151	ug/L	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091621\
Data File : VV022168.D
Acq On : 16 Sep 2021 20:09
Operator : SY/MD
Sample : M3764-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CA8

Quant Time: Sep 17 01:58:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 17 01:54:12 2021
Response via : Initial Calibration

