

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
 Data File : VV027890.D
 Acq On : 16 Sep 2022 18:52
 Operator : SY/MD
 Sample : N4672-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EB4

Quant Time: Sep 16 22:10:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 16 22:04:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	129551	5.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	131096	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	66366	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	32461	3.801	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	76.000%	
7) Chloroethane-d5	1.568	69	36480	5.036	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	100.800%	
11) 1,1-Dichloroethene-d2	2.108	63	49793	2.748	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	55.000%#	
20) 2-Butanone-d5	3.892	46	93001	49.141	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	98.280%	
24) Chloroform-d	4.346	84	98005	5.605	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	112.000%	
26) 1,2-Dichloroethane-d4	5.031	65	51349	5.514	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.200%	
32) Benzene-d6	5.047	84	167629	4.408	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
36) 1,2-Dichloropropane-d6	6.066	67	48138	4.323	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	86.400%	
41) Toluene-d8	7.313	98	135475	4.236	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.800%	
43) trans-1,3-Dichloroprop...	7.622	79	15583	4.171	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.400%	
46) 2-Hexanone-d5	8.088	63	67242	48.425	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	96.840%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	29902	4.182	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	83.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55699	5.109	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
Data File : VV027890.D
Acq On : 16 Sep 2022 18:52
Operator : SY/MD
Sample : N4672-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EB4

Quant Time: Sep 16 22:10:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 16 22:04:36 2022
Response via : Initial Calibration

