

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091622\
 Data File : VW027891.D
 Acq On : 16 Sep 2022 19:16
 Operator : SY/MD
 Sample : N4672-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EB8

Quant Time: Sep 16 22:10:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 16 22:04:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	121928	5.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	123092	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62803	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	29535	3.675	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.400%
7) Chloroethane-d5	1.568	69	33748	4.950	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.108	63	48954	2.870	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.400%#
20) 2-Butanone-d5	3.889	46	88082	49.452	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.900%
24) Chloroform-d	4.349	84	92550	5.624	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.400%
26) 1,2-Dichloroethane-d4	5.030	65	47763	5.450	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.000%
32) Benzene-d6	5.047	84	154967	4.340	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.066	67	44361	4.243	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.800%
41) Toluene-d8	7.313	98	125518	4.180	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	7.625	79	15521	4.425	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.600%
46) 2-Hexanone-d5	8.088	63	64912	49.787	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.580%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	32805	4.886	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	51903	5.031	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
Data File : VV027891.D
Acq On : 16 Sep 2022 19:16
Operator : SY/MD
Sample : N4672-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EB8

Quant Time: Sep 16 22:10:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 16 22:04:36 2022
Response via : Initial Calibration

