

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091622\
 Data File : VW027895.D
 Acq On : 16 Sep 2022 20:53
 Operator : SY/MD
 Sample : N4672-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0ED0

Quant Time: Sep 16 22:11:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 16 22:04:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	132141	5.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	132914	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68182	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31193	3.581	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.600%
7) Chloroethane-d5	1.568	69	37542	5.081	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.600%
11) 1,1-Dichloroethene-d2	2.108	63	52595	2.846	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.000%#
20) 2-Butanone-d5	3.892	46	96592	50.038	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.080%
24) Chloroform-d	4.346	84	99172	5.561	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.200%
26) 1,2-Dichloroethane-d4	5.034	65	52204	5.496	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.000%
32) Benzene-d6	5.047	84	169417	4.394	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.800%
36) 1,2-Dichloropropane-d6	6.069	67	48906	4.332	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.600%
41) Toluene-d8	7.317	98	136251	4.202	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	7.625	79	16233	4.286	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.800%
46) 2-Hexanone-d5	8.088	63	69951	49.687	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.380%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	35326	4.873	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	58093	5.187	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
Data File : VV027895.D
Acq On : 16 Sep 2022 20:53
Operator : SY/MD
Sample : N4672-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0ED0

Quant Time: Sep 16 22:11:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 16 22:04:36 2022
Response via : Initial Calibration

