

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091622\
 Data File : VW027877.D
 Acq On : 16 Sep 2022 13:13
 Operator : SY/MD
 Sample : N4618-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E57

Quant Time: Sep 16 22:07:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 16 22:04:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	138038	5.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	137584	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	65584	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33395	3.670	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.400%
7) Chloroethane-d5	1.568	69	36188	4.688	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.800%
11) 1,1-Dichloroethene-d2	2.108	63	47773	2.474	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.400%#
20) 2-Butanone-d5	3.895	46	76979	38.174	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.340%
24) Chloroform-d	4.346	84	94148	5.053	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.000%
26) 1,2-Dichloroethane-d4	5.030	65	47794	4.817	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.047	84	165773	4.153	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.066	67	46704	3.997	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.000%
41) Toluene-d8	7.313	98	133215	3.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	7.622	79	14707	3.751	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.000%
46) 2-Hexanone-d5	8.088	63	56200	38.564	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.120%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	30943	4.123	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	49686	4.612	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
Data File : VV027877.D
Acq On : 16 Sep 2022 13:13
Operator : SY/MD
Sample : N4618-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E57

Quant Time: Sep 16 22:07:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 16 22:04:36 2022
Response via : Initial Calibration

