

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091622\
 Data File : VW027880.D
 Acq On : 16 Sep 2022 14:26
 Operator : SY/MD
 Sample : N4618-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E66

Quant Time: Sep 16 22:08:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 16 22:04:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	136007	5.000	ug/L	# 0.00
28) Chlorobenzene-d5	8.850	117	134187	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	66150	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31181	3.478	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.600%
7) Chloroethane-d5	1.568	69	35784	4.705	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.108	63	44848	2.357	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	47.200%#
20) 2-Butanone-d5	3.895	46	85240	42.902	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.800%
24) Chloroform-d	4.346	84	93387	5.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	5.030	65	50166	5.132	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.600%
32) Benzene-d6	5.047	84	163028	4.188	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
36) 1,2-Dichloropropane-d6	6.066	67	47452	4.163	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.200%
41) Toluene-d8	7.313	98	131670	4.022	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	7.625	79	15952	4.172	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.400%
46) 2-Hexanone-d5	8.088	63	68877	48.460	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.920%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	34626	4.731	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	53969	4.966	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
Data File : VV027880.D
Acq On : 16 Sep 2022 14:26
Operator : SY/MD
Sample : N4618-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E66

Quant Time: Sep 16 22:08:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 16 22:04:36 2022
Response via : Initial Calibration

