



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091622\
 Data File : VV027870.D
 Acq On : 16 Sep 2022 09:40
 Operator : SY/MD
 Sample : BFB371
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB371

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR090622WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Mon Sep 19 16:41:57 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2796

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	11563	PASS
75	95	30	80	46.2	33512	PASS
95	95	100	100	100.0	72605	PASS
96	95	5	9	6.7	4865	PASS
173	174	0.00	2	0.3	178	PASS
174	95	50	120	95.4	69251	PASS
175	174	5	9	7.2	4970	PASS
176	174	95	105	98.3	68072	PASS
177	176	5	10	6.1	4149	PASS

Scan 2796 (10.030 min): VV027870.D\data.ms
BFB371

m/z	Abundance
40.10	165.0
44.00	154.0

Instrument :
MSVOA_V
ClientSampleId :
BFB371

m/z	Abundance
35.90	464.0
37.10	2685.0
38.00	2751.0
39.00	980.0
40.10	196.0
43.90	387.0
45.10	425.0
47.10	555.0
48.00	367.0
49.00	2459.0
50.00	11558.0
51.10	3484.0
52.00	235.0
54.90	169.0
56.00	1134.0
57.00	1510.0
58.10	103.0
59.90	479.0
60.20	526.0
61.00	3293.0
62.00	3123.0
63.00	2901.0
64.00	418.0
65.00	465.0
66.80	236.0
68.00	6915.0
69.00	7066.0
69.90	437.0
72.00	293.0
73.10	2793.0
74.00	11650.0
75.00	32888.0
76.10	2795.0
77.10	400.0
78.90	2467.0
80.00	617.0
80.90	2538.0
82.00	628.0
86.00	104.0
86.90	2818.0
87.90	2490.0
90.80	414.0
92.00	2147.0
93.00	3078.0
94.00	7844.0
95.00	71400.0
96.10	4752.0
104.00	281.0
104.90	106.0
106.00	286.0
114.90	130.0
115.90	317.0
116.90	652.0
118.00	240.0
118.90	542.0
127.90	335.0
128.90	173.0
129.90	228.0
136.80	160.0
140.90	756.0
142.10	126.0
142.90	1226.0
144.90	643.0
148.10	190.0
149.60	110.0
154.90	330.0
156.80	146.0
158.80	106.0
161.00	106.0
171.00	257.0
172.20	424.0
172.80	533.0
174.00	67200.0
175.00	4753.0
176.00	65576.0
177.10	4061.0
178.00	122.0

Instrument :
MSVOA_V
ClientSampleId :
BFB371

m/z	Abundance
35.90	552.0
37.10	2660.0
38.00	2807.0
39.00	1045.0
40.10	211.0
44.00	462.0
44.90	494.0
47.00	511.0
47.90	345.0
49.00	2550.0
50.00	11950.0
51.00	3697.0
52.00	189.0
52.90	110.0
55.10	189.0
56.00	1083.0
57.00	1548.0
58.10	117.0
59.90	510.0
60.10	511.0
61.00	3429.0
62.00	3137.0
63.00	3126.0
64.00	428.0
65.00	536.0
66.80	258.0
68.00	7371.0
69.00	7107.0
70.00	438.0
72.00	366.0
73.00	2905.0
74.10	12003.0
75.00	34736.0
76.10	2749.0
77.10	411.0
78.90	2599.0
80.00	668.0
80.90	2695.0
82.00	585.0
86.90	2964.0
87.90	2624.0
90.80	384.0
92.00	2269.0
93.00	3183.0
94.00	8272.0
95.00	75152.0
96.10	5036.0
97.00	118.0
103.90	320.0
105.80	291.0
114.90	115.0
116.00	337.0
117.00	626.0
118.00	273.0
118.90	583.0
127.90	340.0
128.90	196.0
129.80	281.0
130.80	147.0
136.80	138.0
140.90	818.0
142.10	143.0
142.90	1339.0
144.00	105.0
144.90	655.0
148.00	202.0
148.80	104.0
154.90	307.0
156.90	151.0
158.80	115.0
161.00	109.0
171.00	290.0
172.10	451.0
174.00	71744.0
175.00	5188.0
176.00	70672.0
177.10	4271.0
178.10	125.0

Instrument :
MSVOA_V
ClientSampleId :
BFB371

m/z	Abundance
36.00	549.0
37.00	2502.0
38.00	2406.0
39.00	1016.0
40.10	195.0
44.00	504.0
44.90	503.0
46.90	505.0
47.90	319.0
49.00	2510.0
50.00	11180.0
51.00	3687.0
52.10	181.0
52.90	157.0
55.00	182.0
55.90	929.0
57.00	1483.0
58.00	124.0
59.90	529.0
61.00	3219.0
62.00	2938.0
63.00	2901.0
64.00	430.0
65.10	502.0
66.80	262.0
68.00	6763.0
69.00	6359.0
70.00	407.0
72.00	398.0
73.00	2752.0
74.00	11267.0
75.00	32912.0
76.10	2593.0
77.00	362.0
77.80	310.0
78.90	2442.0
80.00	627.0
81.00	2635.0
82.00	503.0
87.00	2704.0
87.90	2556.0
90.80	293.0
92.00	2104.0
93.00	2971.0
94.00	7787.0
95.00	71264.0
96.10	4807.0
97.00	144.0
103.90	294.0
105.80	333.0
116.00	301.0
116.90	555.0
117.90	304.0
118.90	543.0
127.90	318.0
128.90	172.0
129.80	306.0
130.80	176.0
134.90	105.0
136.70	103.0
137.10	130.0
141.00	795.0
142.00	117.0
142.90	1303.0
144.90	580.0
147.90	215.0
148.80	119.0
149.90	104.0
155.00	247.0
156.90	153.0
158.80	120.0
161.00	112.0
171.00	296.0
172.10	452.0
174.00	68808.0
175.00	4970.0
176.00	67968.0
177.00	4114.0
178.10	112.0

Instrument :
MSVOA_V
ClientSampleId :
BFB371