

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VW091625\
 Data File : VW039098.D
 Acq On : 16 Sep 2025 12:01
 Operator : SY/MD
 Sample : VIBLK221
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK221

Quant Time: Sep 17 02:21:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091525WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 17 02:19:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	364274	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	375727	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.175	152	136096	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.291	65	134096	4.668	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.400%	
7) Chloroethane-d5	1.545	69	119241	4.789	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.800%	
11) 1,1-Dichloroethene-d2	2.072	65	62936	4.598	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	92.000%	
20) 2-Butanone-d5	3.805	46	171571	40.207	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	80.420%	
24) Chloroform-d	4.256	84	283062	5.073	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.400%	
26) 1,2-Dichloroethane-d4	4.947	65	121934	5.189	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.800%	
32) Benzene-d6	4.963	84	488674	4.753	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.000%	
36) 1,2-Dichloropropane-d6	5.989	67	161075	5.242	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.800%	
41) Toluene-d8	7.239	98	372371	4.046	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.000%	
43) trans-1,3-Dichloroprop...	7.558	79	47067	4.187	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.800%	
46) 2-Hexanone-d5	8.017	63	157372	43.247	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	86.500%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	110944	5.057	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.200%	
66) 1,2-Dichlorobenzene-d4	11.548	152	120831	5.516	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	110.400%	
Target Compounds						Qvalue
16) Methylene chloride	2.461	84	5061	0.150	ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091625\
Data File : VV039098.D
Acq On : 16 Sep 2025 12:01
Operator : SY/MD
Sample : VIBLK221
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK221

Quant Time: Sep 17 02:21:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091525WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 17 02:19:46 2025
Response via : Initial Calibration

