

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091718\
 Data File : VV007587.D
 Acq On : 17 Sep 2018 14:28
 Operator : SY/MD
 Sample : J4888-15
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 18 04:13:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 18 03:38:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	116458	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	103229	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	55250	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22971	3.78	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.60%
7) Chloroethane-d5	1.59	69	22689	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.14	63	42124	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	3.97	46	36777	43.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.68%
24) Chloroform-d	4.41	84	56102	4.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.00%
26) 1,2-Dichloroethane-d4	5.09	65	28858	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
32) Benzene-d6	5.10	84	110797	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
36) 1,2-Dichloropropane-d6	6.13	67	32487	4.65	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.00%
41) Toluene-d8	7.36	98	113493	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloropropene-	7.67	79	13528	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
46) 2-Hexanone-d5	8.15	63	19550	37.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.82%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21595	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	49434	4.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091718\
Data File : VV007587.D
Acq On : 17 Sep 2018 14:28
Operator : SY/MD
Sample : J4888-15
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 18 04:13:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 18 03:38:22 2018
Response via : Initial Calibration

