

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091718\
 Data File : VV007588.D
 Acq On : 17 Sep 2018 14:55
 Operator : SY/MD
 Sample : J4918-13
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0KY3

Quant Time: Sep 18 04:18:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 18 03:38:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	116205	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	107169	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	55913	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22813	3.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.20%
7) Chloroethane-d5	1.59	69	21471	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.40%
11) 1,1-Dichloroethene-d2	2.13	63	41548	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	3.97	46	40309	47.61	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.22%
24) Chloroform-d	4.41	84	54272	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.09	65	28420	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
32) Benzene-d6	5.10	84	105906	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
36) 1,2-Dichloropropane-d6	6.13	67	31806	4.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.80%
41) Toluene-d8	7.36	98	109901	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
43) trans-1,3-Dichloropropene-	7.67	79	13373	3.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.80%
46) 2-Hexanone-d5	8.15	63	23699	44.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.54%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20608	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	47691	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	3079	0.45 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091718\
Data File : VV007588.D
Acq On : 17 Sep 2018 14:55
Operator : SY/MD
Sample : J4918-13
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KY3

Quant Time: Sep 18 04:18:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 18 03:38:22 2018
Response via : Initial Calibration

