

Data File : VV007593.D
Acq On : 17 Sep 2018 17:58
Operator : SY/MD
Sample : J4864-07
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 18 04:36:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 18 03:38:22 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	118198	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	107975	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56362	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	22742	3.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.60%
7) Chloroethane-d5	1.59	69	23240	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.80%
11) 1,1-Dichloroethene-d2	2.14	63	41980	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	3.97	46	52885	61.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.82%
24) Chloroform-d	4.41	84	55940	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
26) 1,2-Dichloroethane-d4	5.09	65	29994	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.10	84	111041	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
36) 1,2-Dichloropropane-d6	6.13	67	32993	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.40%
41) Toluene-d8	7.36	98	113081	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
43) trans-1,3-Dichloropropene-	7.67	79	12797	3.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.80%
46) 2-Hexanone-d5	8.15	63	20484	37.98	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.96%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21465	4.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	48988	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007593.D
Acq On : 17 Sep 2018 17:58
Operator : SY/MD
Sample : J4864-07
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 18 04:36:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091318WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 18 03:38:22 2018
Response via : Initial Calibration

