

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091720\
 Data File : VV018295.D
 Acq On : 17 Sep 2020 21:45
 Operator : SY/MD
 Sample : VV0917WBL02
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB93

Quant Time: Sep 18 08:06:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 18 08:05:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	278779	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	275490	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	134484	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87847	38.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.34%
7) Chloroethane-d5	1.58	69	77232	41.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.08%
11) 1,1-Dichloroethene-d2	2.12	63	133175	31.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.92%
21) 2-Butanone-d5	3.92	46	126856	91.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.22%
24) Chloroform-d	4.38	84	171764	43.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.14%
26) 1,2-Dichloroethane-d4	5.06	65	113853	45.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.12%
32) Benzene-d6	5.08	84	339549	43.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.10%
36) 1,2-Dichloropropane-d6	6.09	67	110806	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.44%
41) Toluene-d8	7.34	98	300809	41.85	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.70%
43) trans-1,3-Dichloropropene-	7.64	79	52912	41.55	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.10%
47) 2-Hexanone-d5	8.11	63	91371	96.26	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.26%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	148122	47.10	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.20%
64) 1,2-Dichlorobenzene-d4	11.65	152	131264	47.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.06%

Target Compounds


					Ovalue
10) 1,1,2-Trichloro-1,2,2-trif	2.12	101	1251	0.665 ug/L #	19
35) Methylcyclohexane	6.09	83	26595	8.232 ug/L #	18
37) 1,2-Dichloropropane	6.10	63	12007	5.753 ug/L #	87
39) cis-1,3-Dichloropropene	7.01	75	2892	0.899 ug/L #	62
59) 1,2,3-Trichloropropane	11.27	75	15416	6.280 ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091720\
Data File : VV018295.D
Acq On : 17 Sep 2020 21:45
Operator : SY/MD
Sample : VV0917WBL02
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB93

Quant Time: Sep 18 08:06:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 18 08:05:39 2020
Response via : Initial Calibration





#10

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 0.665 ug/L

RT: 2.12 min Scan# 321

Delta R.T. -0.01 min

Lab File: VV018295.D

Acq: 17 Sep 2020 21:45

Instrument :

MSVOA_V

ClientSampled :

BFB93


Tot Ion:101 Resp: 1251

Ion Ratio Lower Upper

101 100

85 0.0 33.5 50.3#




151 0.0 62.2 93.2#



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VV0

Ion 151.00 (150.70 to 151.70): V



#35

Methylcyclohexane

Concen: 8.232 ug/L

RT: 6.09 min Scan# 1556

Delta R.T. -0.06 min

Lab File: VV018295.D

Acq: 17 Sep 2020 21:45


Tgt Ion: 83 Resp: 26595

Ion Ratio Lower Upper

83 100

55 0.5 63.4 95.2#



98 1.5 38.8 58.2#




Abundance Ion 83.00 (82.70 to 83.70): VV0

Ion 55.00 (54.70 to 55.70): VV0

Ion 98.00 (97.70 to 98.70): VV0









#37
 1,2-Dichloropropane
 Concen: 5.753 ug/L
 RT: 6.10 min Scan# 1557
 Delta R.T. -0.10 min
 Lab File: VV018295.D
 Acq: 17 Sep 2020 21:45




Instrument :
 MSVOA_V
 ClientSampleId :
 BFB93


Tot Ion: 63 Resp: 12007
 Ion Ratio Lower Upper
 63 100
 112 0.0 3.4 5.2#



#39
 cis-1,3-Dichloropropene
 Concen: 0.899 ug/L
 RT: 7.01 min Scan# 1840
 Delta R.T. -0.04 min
 Lab File: VV018295.D
 Acq: 17 Sep 2020 21:45

Tgt Ion: 75 Resp: 2892
 Ion Ratio Lower Upper
 75 100
 77 53.0 22.3 41.3#





#59

1,2,3-Trichloropropane

Concen: 6.280 ug/L

RT: 11.27 min Scan# 3166

Delta R.T. 0.97 min

Lab File: VV018295.D

Acq: 17 Sep 2020 21:45

Instrument :

MSVOA_V

ClientSampleId :

BFB93

Tot Ion: 75 Resp: 15416

Ion Ratio Lower Upper

75 100

77 37.5 25.8 38.6

