

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091720\
 Data File : VV018292.D
 Acq On : 17 Sep 2020 20:15
 Operator : SY/MD
 Sample : L4043-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AE3

Quant Time: Sep 18 08:01:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 18 06:17:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	288439	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	287741	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	137125	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	79821	33.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.92%
7) Chloroethane-d5	1.58	69	71141	36.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.98%
11) 1,1-Dichloroethene-d2	2.12	63	122742	28.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.94%#
21) 2-Butanone-d5	3.92	46	122729	85.30	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.30%
24) Chloroform-d	4.38	84	164391	40.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.62%
26) 1,2-Dichloroethane-d4	5.06	65	110056	42.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.20%
32) Benzene-d6	5.08	84	324478	39.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.78%
36) 1,2-Dichloropropane-d6	6.09	67	105900	40.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.92%
41) Toluene-d8	7.34	98	288904	38.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.96%#
43) trans-1,3-Dichloropropene-	7.64	79	51066	38.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.80%
47) 2-Hexanone-d5	8.11	63	85897	86.64	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.64%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	142277	43.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	125299	44.49	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091720\
Data File : VV018292.D
Acq On : 17 Sep 2020 20:15
Operator : SY/MD
Sample : L4043-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AE3

Quant Time: Sep 18 08:01:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 18 06:17:41 2020
Response via : Initial Calibration

