

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091720\
 Data File : VV018293.D
 Acq On : 17 Sep 2020 20:37
 Operator : SY/MD
 Sample : L4043-10
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AE4

Quant Time: Sep 18 08:04:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 18 06:17:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	283834	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	283952	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	134427	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	76726	33.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	66.34%
7) Chloroethane-d5	1.58	69	68407	36.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.28%
11) 1,1-Dichloroethene-d2	2.12	63	117703	27.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.48%#
21) 2-Butanone-d5	3.92	46	115816	81.80	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.80%
24) Chloroform-d	4.38	84	157899	39.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.68%
26) 1,2-Dichloroethane-d4	5.06	65	107555	41.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.62%
32) Benzene-d6	5.08	84	316283	38.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.82%
36) 1,2-Dichloropropane-d6	6.09	67	104040	40.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.56%
41) Toluene-d8	7.34	98	279125	37.68	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.36%#
43) trans-1,3-Dichloropropene-	7.64	79	49477	37.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.40%
47) 2-Hexanone-d5	8.11	63	81619	83.43	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.43%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	137778	42.50	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	121894	44.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.30%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091720\
Data File : VV018293.D
Acq On : 17 Sep 2020 20:37
Operator : SY/MD
Sample : L4043-10
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AE4

Quant Time: Sep 18 08:04:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 18 06:17:41 2020
Response via : Initial Calibration

