

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091720\
 Data File : VV018299.D
 Acq On : 17 Sep 2020 23:16
 Operator : SY/MD
 Sample : L4043-14
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF9

Quant Time: Sep 18 08:23:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 18 08:05:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	278527	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	276963	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	131606	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	76216	33.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	67.16%
7) Chloroethane-d5	1.58	69	69242	37.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.56%
11) 1,1-Dichloroethene-d2	2.12	63	117054	28.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.22%#
21) 2-Butanone-d5	3.92	46	118006	84.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.94%
24) Chloroform-d	4.38	84	160152	40.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.34%
26) 1,2-Dichloroethane-d4	5.06	65	106716	42.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.54%
32) Benzene-d6	5.08	84	312526	39.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.84%
36) 1,2-Dichloropropane-d6	6.10	67	103553	41.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.20%
41) Toluene-d8	7.34	98	278174	38.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.00%#
43) trans-1,3-Dichloropropene-	7.64	79	48652	38.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.00%
47) 2-Hexanone-d5	8.11	63	81552	85.46	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.46%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	137958	43.63	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.26%
64) 1,2-Dichlorobenzene-d4	11.65	152	121450	44.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.86%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091720\
Data File : VV018299.D
Acq On : 17 Sep 2020 23:16
Operator : SY/MD
Sample : L4043-14
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF9

Quant Time: Sep 18 08:23:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 18 08:05:39 2020
Response via : Initial Calibration

