

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091720\
 Data File : VV018275.D
 Acq On : 17 Sep 2020 13:50
 Operator : SY/MD
 Sample : L4042-21
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 18 06:55:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 18 06:17:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	269863	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	268417	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	128740	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	104607	47.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.12%
7) Chloroethane-d5	1.58	69	86730	48.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.38%
11) 1,1-Dichloroethene-d2	2.12	63	151205	37.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.96%
21) 2-Butanone-d5	3.92	46	149251	110.87	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.87%
24) Chloroform-d	4.38	84	195220	51.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.32%
26) 1,2-Dichloroethane-d4	5.06	65	131709	53.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.70%
32) Benzene-d6	5.08	84	389719	50.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.44%
36) 1,2-Dichloropropane-d6	6.09	67	127191	52.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.18%
41) Toluene-d8	7.34	98	346450	49.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.94%
43) trans-1,3-Dichloropropene-	7.64	79	62360	50.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.52%
47) 2-Hexanone-d5	8.11	63	105549	114.13	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.13%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	168023	54.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.66%
64) 1,2-Dichlorobenzene-d4	11.65	152	147454	55.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.54%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091720\
Data File : VV018275.D
Acq On : 17 Sep 2020 13:50
Operator : SY/MD
Sample : L4042-21
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 18 06:55:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 18 06:17:41 2020
Response via : Initial Calibration

