

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091720\  
 Data File : VV018288.D  
 Acq On : 17 Sep 2020 18:44  
 Operator : SY/MD  
 Sample : L4066-11  
 Misc : 5.0mL/MSVOA V/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 BFWY9

Quant Time: Sep 18 07:31:06 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVLM090920WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Fri Sep 18 06:17:41 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.64  | 114  | 284586   | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.87  | 117  | 281975   | 50.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.27 | 152  | 137001   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |         |       |          |          |      |        |
|--------------------------------|---------|-------|----------|----------|------|--------|
| 4) Vinyl Chloride-d3           | 1.32    | 65    | 86419    | 37.26    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 60 - 135 | Recovery | =    | 74.52% |
| 7) Chloroethane-d5             | 1.58    | 69    | 75383    | 39.72    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 130 | Recovery | =    | 79.44% |
| 11) 1,1-Dichloroethene-d2      | 2.12    | 63    | 128597   | 30.23    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 60.46% |
| 21) 2-Butanone-d5              | 3.92    | 46    | 122399   | 86.22    | ug/L | 0.00   |
| Spiked Amount                  | 100.000 | Range | 40 - 130 | Recovery | =    | 86.22% |
| 24) Chloroform-d               | 4.38    | 84    | 171477   | 42.61    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 85.22% |
| 26) 1,2-Dichloroethane-d4      | 5.06    | 65    | 115464   | 44.76    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 89.52% |
| 32) Benzene-d6                 | 5.08    | 84    | 336573   | 41.70    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 83.40% |
| 36) 1,2-Dichloropropane-d6     | 6.09    | 67    | 111320   | 43.40    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 70 - 120 | Recovery | =    | 86.80% |
| 41) Toluene-d8                 | 7.34    | 98    | 299664   | 40.73    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 81.46% |
| 43) trans-1,3-Dichloropropene- | 7.64    | 79    | 52506    | 40.29    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 80.58% |
| 47) 2-Hexanone-d5              | 8.11    | 63    | 90163    | 92.80    | ug/L | 0.00   |
| Spiked Amount                  | 100.000 | Range | 45 - 130 | Recovery | =    | 92.80% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.24   | 84    | 147963   | 45.96    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 65 - 120 | Recovery | =    | 91.92% |
| 64) 1,2-Dichlorobenzene-d4     | 11.65   | 152   | 130271   | 46.30    | ug/L | 0.00   |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 92.60% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091720\  
Data File : VV018288.D  
Acq On : 17 Sep 2020 18:44  
Operator : SY/MD  
Sample : L4066-11  
Misc : 5.0mL/MSVOA V/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
BFWY9

Quant Time: Sep 18 07:31:06 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVLM090920WMA.M  
Quant Title : VOC Analysis  
QLast Update : Fri Sep 18 06:17:41 2020  
Response via : Initial Calibration

