

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091721\
 Data File : VV022183.D
 Acq On : 17 Sep 2021 13:57
 Operator : SY/MD
 Sample : VIBLK208
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK208

Quant Time: Sep 18 02:57:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 18 02:49:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	228645	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	220084	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	92470	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	55404	34.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.980%
7) Chloroethane-d5	1.561	69	59333	47.984	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.960%
11) 1,1-Dichloroethene-d2	2.104	63	86622	33.079	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.160%
21) 2-Butanone-d5	3.876	46	155017	142.752	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	142.750%#
24) Chloroform-d	4.349	84	162897	56.421	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.840%
26) 1,2-Dichloroethane-d4	5.034	65	110801	65.154	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	130.300%#
32) Benzene-d6	5.053	84	314319	52.655	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.300%
36) 1,2-Dichloropropane-d6	6.072	67	108981	58.052	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	116.100%
41) Toluene-d8	7.320	98	259634	46.075	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.160%
43) trans-1,3-Dichloroprop...	7.625	79	44772	50.107	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.220%
47) 2-Hexanone-d5	8.088	63	89670	124.487	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	124.490%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	145839	60.657	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	121.320%#
66) 1,2-Dichlorobenzene-d4	11.625	152	101348	55.751	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	111.500%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.510	84	2411	1.364	ug/L	90
46) Tetrachloroethene	7.979	164	16030	11.366	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091721\
Data File : VV022183.D
Acq On : 17 Sep 2021 13:57
Operator : SY/MD
Sample : VIBLK208
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK208

Quant Time: Sep 18 02:57:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 18 02:49:48 2021
Response via : Initial Calibration

