

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091721\
 Data File : VW022187.D
 Acq On : 17 Sep 2021 15:52
 Operator : SY/MD
 Sample : M3676-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CC9

Quant Time: Sep 18 02:58:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 18 02:49:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	288146	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	273359	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	123299	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	73477	36.822	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.640%
7) Chloroethane-d5	1.568	69	68589	44.015	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.040%
11) 1,1-Dichloroethene-d2	2.105	63	117794	35.694	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.380%
21) 2-Butanone-d5	3.899	46	153590	112.231	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.230%
24) Chloroform-d	4.352	84	181433	49.865	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.720%
26) 1,2-Dichloroethane-d4	5.037	65	116096	54.171	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.340%
32) Benzene-d6	5.053	84	366901	49.485	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.960%
36) 1,2-Dichloropropane-d6	6.072	67	119162	51.104	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.200%
41) Toluene-d8	7.320	98	324645	46.384	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.760%
43) trans-1,3-Dichloroprop...	7.625	79	47194	42.524	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.040%
47) 2-Hexanone-d5	8.092	63	96813	108.209	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.210%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	153063	51.255	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.500%
66) 1,2-Dichlorobenzene-d4	11.625	152	124945	51.547	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.100%
Target Compounds						
					Qvalue	
13) Acetone	2.185	43	14287	12.460	ug/L	# 74
34) Trichloroethene	5.924	95	4864	2.159	ug/L	89
46) Tetrachloroethene	7.982	164	2109	1.204	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091721\
Data File : VV022187.D
Acq On : 17 Sep 2021 15:52
Operator : SY/MD
Sample : M3676-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CC9

Quant Time: Sep 18 02:58:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 18 02:49:48 2021
Response via : Initial Calibration

