

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW091721\
 Data File : VW022199.D
 Acq On : 17 Sep 2021 21:06
 Operator : SY/MD
 Sample : M3676-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0CE6

Quant Time: Sep 18 04:43:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 18 04:25:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	299060	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	281598	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	122889	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	70670	34.123	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.240%
7) Chloroethane-d5	1.568	69	67498	41.734	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.460%
11) 1,1-Dichloroethene-d2	2.105	63	113090	33.018	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.040%
21) 2-Butanone-d5	3.892	46	154531	108.798	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.800%
24) Chloroform-d	4.349	84	179669	47.578	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.160%
26) 1,2-Dichloroethane-d4	5.034	65	113723	51.127	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.260%
32) Benzene-d6	5.050	84	359612	47.083	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.160%
36) 1,2-Dichloropropane-d6	6.072	67	118963	49.526	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.060%
41) Toluene-d8	7.317	98	318240	44.139	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.280%
43) trans-1,3-Dichloroprop...	7.625	79	46034	40.265	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.540%
47) 2-Hexanone-d5	8.092	63	92587	100.458	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.460%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	151857	49.363	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.720%
66) 1,2-Dichlorobenzene-d4	11.628	152	121637	50.349	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091721\
Data File : VV022199.D
Acq On : 17 Sep 2021 21:06
Operator : SY/MD
Sample : M3676-18
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0CE6

Quant Time: Sep 18 04:43:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 18 04:25:40 2021
Response via : Initial Calibration

