

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091724\
 Data File : VV037307.D
 Acq On : 17 Sep 2024 10:43
 Operator : SY/MD
 Sample : P3391-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT3-2024-06

Quant Time: Sep 18 05:37:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 14 03:38:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	12229	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	11983	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	5880	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	7320	0.559	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.000%
4) 1,2-Dichloroethane-d4	4.969	65	4395	0.589	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	118.000%
7) 1,2-Dichloropropane-d6	5.996	67	5489	0.552	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.000%
8) Toluene-d8	7.254	98	14170	0.485	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	4131	0.555	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	110.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	521	0.044	ug/L #	57
6) Trichloroethene	5.852	95	335	0.036	ug/L	94
9) 1,2-Dibromoethane	8.321	107	193	0.034	ug/L #	88
12) 1,2,3-Trichloropropane	10.214	75	182	0.038	ug/L	92
13) 1,2-Dibromo-3-chloropr...	12.374	75	36	0.042	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091724\
Data File : VV037307.D
Acq On : 17 Sep 2024 10:43
Operator : SY/MD
Sample : P3391-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT3-2024-06

Quant Time: Sep 18 05:37:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 14 03:38:32 2024
Response via : Initial Calibration

