

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091724\
 Data File : VV037311.D
 Acq On : 17 Sep 2024 12:43
 Operator : SY/MD
 Sample : P3967-18DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AP9DL

Quant Time: Sep 18 05:38:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 14 03:38:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	14116	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	13913	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	7163	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	7227	0.478	ug/L	0.00
Spiked Amount 0.500	Range	40 - 130	Recovery	=	96.000%	
4) 1,2-Dichloroethane-d4	4.969	65	4646	0.540	ug/L	0.00
Spiked Amount 0.500	Range	70 - 130	Recovery	=	108.000%	
7) 1,2-Dichloropropane-d6	5.995	67	6020	0.521	ug/L	0.00
Spiked Amount 0.500	Range	60 - 140	Recovery	=	104.000%	
8) Toluene-d8	7.254	98	15914	0.470	ug/L	0.00
Spiked Amount 0.500	Range	70 - 130	Recovery	=	94.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	4851	0.561	ug/L	0.00
Spiked Amount 0.500	Range	65 - 120	Recovery	=	112.000%	
Target Compounds					Qvalue	
6) Trichloroethene	5.852	95	6550	0.602	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091724\
Data File : VV037311.D
Acq On : 17 Sep 2024 12:43
Operator : SY/MD
Sample : P3967-18DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AP9DL

Quant Time: Sep 18 05:38:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 14 03:38:32 2024
Response via : Initial Calibration

