

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091724\
 Data File : VV037313.D
 Acq On : 17 Sep 2024 13:32
 Operator : SY/MD
 Sample : P3990-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AS1

Quant Time: Sep 18 05:38:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 14 03:38:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	14092	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	13953	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	7048	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	8432	0.559	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 112.000%	
4) 1,2-Dichloroethane-d4	4.969	65	5094	0.593	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 118.000%	
7) 1,2-Dichloropropane-d6	5.995	67	6507	0.562	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 112.000%	
8) Toluene-d8	7.253	98	17254	0.508	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 102.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	4802	0.554	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 110.000%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091724\
Data File : VV037313.D
Acq On : 17 Sep 2024 13:32
Operator : SY/MD
Sample : P3990-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AS1

Quant Time: Sep 18 05:38:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 14 03:38:32 2024
Response via : Initial Calibration

