

Data File : VV007624.D
Acq On : 19 Sep 2018 01:31
Operator : SY/MD
Sample : J4874-02
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A3Z75

Quant Time: Sep 19 04:05:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	109300	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	100587	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	52649	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26200	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.59	69	24312	5.20	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.00%
11) 1,1-Dichloroethene-d2	2.14	63	45560	3.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.40%
20) 2-Butanone-d5	3.97	46	45752	63.68	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.36%
24) Chloroform-d	4.41	84	55210	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
26) 1,2-Dichloroethane-d4	5.09	65	28905	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.10	84	111236	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.60%
36) 1,2-Dichloropropane-d6	6.12	67	32807	5.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.40%
41) Toluene-d8	7.36	98	113456	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
43) trans-1,3-Dichloropropene-	7.67	79	12333	4.54	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.80%
46) 2-Hexanone-d5	8.15	63	22868	47.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.02%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	18088	4.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	48711	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds					Qvalue
25) Chloroform	4.43	83	22357	1.778 ug/L	99
27) 1,2-Dichloroethane	5.19	62	2207	0.307 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007624.D
Acq On : 19 Sep 2018 01:31
Operator : SY/MD
Sample : J4874-02
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A3Z75

Quant Time: Sep 19 04:05:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

