

Data File : VV007625.D
Acq On : 19 Sep 2018 01:58
Operator : SY/MD
Sample : J4874-04
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A3ZB3

Quant Time: Sep 19 04:07:14 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	112502	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	101047	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	52890	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26021	4.85	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.00%
7) Chloroethane-d5	1.59	69	24202	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.60%
11) 1,1-Dichloroethene-d2	2.14	63	46543	3.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.80%
20) 2-Butanone-d5	3.97	46	56952	77.02	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	154.04%#
24) Chloroform-d	4.41	84	55858	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
26) 1,2-Dichloroethane-d4	5.09	65	29477	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.20%
32) Benzene-d6	5.10	84	113241	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.20%
36) 1,2-Dichloropropane-d6	6.13	67	32658	5.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.40%
41) Toluene-d8	7.36	98	113517	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
43) trans-1,3-Dichloropropene-	7.67	79	12430	4.55	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.00%
46) 2-Hexanone-d5	8.15	63	25859	53.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.96%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	19524	4.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	47802	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007625.D
Acq On : 19 Sep 2018 01:58
Operator : SY/MD
Sample : J4874-04
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A3ZB3

Quant Time: Sep 19 04:07:14 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

