

Data File : VV007609.D
Acq On : 18 Sep 2018 18:45
Operator : SY/MD
Sample : J4887-10
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KR4

Quant Time: Sep 19 02:26:20 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	116527	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	105846	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56667	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25750	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.59	69	23188	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.20%
11) 1,1-Dichloroethene-d2	2.14	63	44138	3.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.80%
20) 2-Butanone-d5	3.97	46	56944	74.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	148.68%#
24) Chloroform-d	4.41	84	54026	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.09	65	28430	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.10	84	108976	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.13	67	31825	4.95	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.00%
41) Toluene-d8	7.36	98	110560	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.60%
43) trans-1,3-Dichloropropene-	7.67	79	12740	4.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.20%
46) 2-Hexanone-d5	8.15	63	25103	49.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.12%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	20330	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	46412	4.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.60%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007609.D
Acq On : 18 Sep 2018 18:45
Operator : SY/MD
Sample : J4887-10
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KR4

Quant Time: Sep 19 02:26:20 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

