

Data File : VV007618.D
Acq On : 18 Sep 2018 22:49
Operator : SY/MD
Sample : J4918-11DL 10X
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KY1DL

Quant Time: Sep 19 03:47:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	117314	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	107025	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	56553	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25695	4.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.80%
7) Chloroethane-d5	1.58	69	23848	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.13	63	46757	3.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.20%
20) 2-Butanone-d5	3.97	46	45548	59.07	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.14%
24) Chloroform-d	4.41	84	53458	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
26) 1,2-Dichloroethane-d4	5.09	65	28394	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.00%
32) Benzene-d6	5.10	84	108805	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.13	67	32314	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.36	98	111044	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	7.67	79	12464	4.31	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.20%
46) 2-Hexanone-d5	8.15	63	24732	48.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.58%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	19320	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	45967	4.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.80%


Target Compounds					Qvalue
12) 1,1-Dichloroethene	2.14	96	4654	0.569 ug/L #	1
22) cis-1,2-Dichloroethene	3.96	96	4983	0.625 ug/L	98
34) Trichloroethene	5.96	95	53702	5.890 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV007618.D
Acq On : 18 Sep 2018 22:49
Operator : SY/MD
Sample : J4918-11DL 10X
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0KY1DL

Quant Time: Sep 19 03:47:55 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091818WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 19 01:47:19 2018
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 0.569 ug/L

RT: 2.14 min Scan# 328

Delta R.T. -0.00 min

Lab File: VV007618.D

Acq: 18 Sep 2018 22:49

Instrument :

MSVOA_V

ClientSampleId :

C0KY1DL





Tgt Ion: 96 Resp: 4654

Ion Ratio Lower Upper

96 100

61 193.2 108.1 200.9

63 478.4 76.9 142.9#



#22

cis-1,2-Dichloroethene

Concen: 0.625 ug/L

RT: 3.96 min Scan# 894

Delta R.T. -0.00 min

Lab File: VV007618.D

Acq: 18 Sep 2018 22:49




Tgt Ion: 96 Resp: 4983


Ion Ratio Lower Upper

96 100

61 110.1 78.4 145.6

68 0.0 0.0 0.0





#34

Trichloroethene

Concen: 5.890 ug/L

RT: 5.96 min Scan# 1516

Delta R.T. 0.00 min

Lab File: VV007618.D

Acq: 18 Sep 2018 22:49

Instrument :

MSVOA_V

ClientSampleId :

C0KY1DL

Tgt Ion: 95 Resp: 53702


Ion Ratio Lower Upper

95 100

97 64.0 46.5 86.5

132 106.5 75.5 140.3

130 112.1 78.8 146.4



Abundance Ion 95.00 (94.70 to 95.70): VV0

Ion 96.95 (96.65 to 97.65): VV0

Ion 131.95 (131.65 to 132.65): V

Ion 129.95 (129.65 to 130.65): V

