

Data File : VV012902.D
Acq On : 18 Sep 2019 18:08
Operator : SY/MD
Sample : K4954-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWK25

Quant Time: Sep 19 04:34:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 19 01:01:31 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	369108	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	368532	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	134742	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	146635	4.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.20%
7) Chloroethane-d5	1.58	69	135971	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.13	63	180946	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	4.00	46	329292	43.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.08%
24) Chloroform-d	4.40	84	261988	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.08	65	139086	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.10	84	538467	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
36) 1,2-Dichloropropane-d6	6.12	67	169963	4.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.40%
41) Toluene-d8	7.36	98	422946	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloropropene-	7.67	79	52968	3.96	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.20%
46) 2-Hexanone-d5	8.14	63	221921	39.49	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.98%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	120259	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	145692	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds					Qvalue
3) Chloromethane	1.25	50	8714	0.339 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV012902.D
Acq On : 18 Sep 2019 18:08
Operator : SY/MD
Sample : K4954-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWK25

Quant Time: Sep 19 04:34:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 19 01:01:31 2019
Response via : Initial Calibration

