

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091820\
 Data File : VV018327.D
 Acq On : 18 Sep 2020 17:00
 Operator : SY/MD
 Sample : L4065-16
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AJ4

Quant Time: Sep 19 04:07:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 19 01:35:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	271254	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	272196	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	129544	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	96608	43.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.40%
7) Chloroethane-d5	1.58	69	79867	44.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.30%
11) 1,1-Dichloroethene-d2	2.12	63	135753	33.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.96%
21) 2-Butanone-d5	3.92	46	116148	85.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.84%
24) Chloroform-d	4.38	84	168014	43.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.62%
26) 1,2-Dichloroethane-d4	5.06	65	110835	45.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.16%
32) Benzene-d6	5.08	84	335639	43.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.14%
36) 1,2-Dichloropropane-d6	6.09	67	108363	43.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.54%
41) Toluene-d8	7.34	98	297986	41.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.92%
43) trans-1,3-Dichloropropene-	7.64	79	50956	40.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.00%
47) 2-Hexanone-d5	8.11	63	79441	84.71	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.71%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	134562	43.30	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	124267	46.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.42%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091820\
Data File : VV018327.D
Acq On : 18 Sep 2020 17:00
Operator : SY/MD
Sample : L4065-16
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AJ4

Quant Time: Sep 19 04:07:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 19 01:35:10 2020
Response via : Initial Calibration

