

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091820\
 Data File : VV018347.D
 Acq On : 19 Sep 2020 00:54
 Operator : SY/MD
 Sample : L4080-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AM1

Quant Time: Sep 19 06:13:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 19 05:11:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	272134	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	270655	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	128844	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	93233	42.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.08%
7) Chloroethane-d5	1.58	69	77293	42.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.18%
11) 1,1-Dichloroethene-d2	2.12	63	137038	33.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.38%
21) 2-Butanone-d5	3.91	46	115084	84.78	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.78%
24) Chloroform-d	4.38	84	170783	44.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.76%
26) 1,2-Dichloroethane-d4	5.06	65	113035	45.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.66%
32) Benzene-d6	5.08	84	341624	44.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.18%
36) 1,2-Dichloropropane-d6	6.09	67	108758	44.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.36%
41) Toluene-d8	7.34	98	302614	42.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.72%
43) trans-1,3-Dichloropropene-	7.64	79	50562	40.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.84%
47) 2-Hexanone-d5	8.11	63	75650	81.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.12%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	134166	43.42	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.84%
64) 1,2-Dichlorobenzene-d4	11.65	152	126846	47.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.88%

Target Compounds					Ovalue
25) Chloroform	4.40	83	9116	2.576 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091820\
Data File : VV018347.D
Acq On : 19 Sep 2020 00:54
Operator : SY/MD
Sample : L4080-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AM1

Quant Time: Sep 19 06:13:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 19 05:11:16 2020
Response via : Initial Calibration

