

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091820\
 Data File : VV018339.D
 Acq On : 18 Sep 2020 21:53
 Operator : SY/MD
 Sample : L4080-01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AL3

Quant Time: Sep 19 05:35:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 19 05:11:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	281004	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	280390	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	135302	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	96382	42.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.18%
7) Chloroethane-d5	1.58	69	79772	42.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.14%
11) 1,1-Dichloroethene-d2	2.12	63	138937	33.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.14%
21) 2-Butanone-d5	3.92	46	115766	82.59	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	82.59%
24) Chloroform-d	4.38	84	168835	42.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.98%
26) 1,2-Dichloroethane-d4	5.06	65	113914	44.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.46%
32) Benzene-d6	5.08	84	343736	42.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.64%
36) 1,2-Dichloropropane-d6	6.09	67	109694	43.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.02%
41) Toluene-d8	7.34	98	304094	41.57	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.14%
43) trans-1,3-Dichloropropene-	7.64	79	51307	39.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.18%
47) 2-Hexanone-d5	8.11	63	76366	79.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.05%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	134724	42.09	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.18%
64) 1,2-Dichlorobenzene-d4	11.65	152	125348	45.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.22%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091820\
Data File : VV018339.D
Acq On : 18 Sep 2020 21:53
Operator : SY/MD
Sample : L4080-01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AL3

Quant Time: Sep 19 05:35:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 19 05:11:16 2020
Response via : Initial Calibration

