

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091820\
 Data File : VV018346.D
 Acq On : 19 Sep 2020 00:31
 Operator : SY/MD
 Sample : L4080-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COAMO

Quant Time: Sep 19 06:10:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 19 05:11:16 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	284058	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	281937	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	134567	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	94199	40.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.38%
7) Chloroethane-d5	1.58	69	79973	42.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.44%
11) 1,1-Dichloroethene-d2	2.12	63	137863	32.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.94%
21) 2-Butanone-d5	3.92	46	114638	80.91	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.91%
24) Chloroform-d	4.38	84	169377	42.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.34%
26) 1,2-Dichloroethane-d4	5.06	65	113559	44.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.22%
32) Benzene-d6	5.08	84	346900	42.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.96%
36) 1,2-Dichloropropane-d6	6.09	67	110230	42.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.96%
41) Toluene-d8	7.34	98	305535	41.54	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.08%
43) trans-1,3-Dichloropropene-	7.64	79	51083	39.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.40%
47) 2-Hexanone-d5	8.11	63	75552	77.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.78%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	131552	40.87	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.74%
64) 1,2-Dichlorobenzene-d4	11.65	152	127634	46.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091820\
Data File : VV018346.D
Acq On : 19 Sep 2020 00:31
Operator : SY/MD
Sample : L4080-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AM0

Quant Time: Sep 19 06:10:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM090920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 19 05:11:16 2020
Response via : Initial Calibration

