

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091924\
 Data File : VV037395.D
 Acq On : 19 Sep 2024 21:04
 Operator : SY/MD
 Sample : VV0919WBL02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK244

Quant Time: Sep 20 03:57:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 20 03:11:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	317693	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	322432	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	146759	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	97590	5.236	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.800%
7) Chloroethane-d5	1.529	69	85316	4.390	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.800%
11) 1,1-Dichloroethene-d2	2.053	65	43243	4.493	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.800%
20) 2-Butanone-d5	3.818	46	234438	42.088	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.180%
24) Chloroform-d	4.243	84	201820	4.460	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	4.941	65	96896	4.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
32) Benzene-d6	4.953	84	397935	4.603	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	5.986	67	125766	4.678	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.600%
41) Toluene-d8	7.239	98	342628	4.399	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
43) trans-1,3-Dichloroprop...	7.561	79	32478	3.673	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.400%
46) 2-Hexanone-d5	8.027	63	210308	48.675	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.360%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	102517	4.679	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.600%
66) 1,2-Dichlorobenzene-d4	11.558	152	124722	4.780	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091924\
Data File : VV037395.D
Acq On : 19 Sep 2024 21:04
Operator : SY/MD
Sample : VV0919WBL02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK244

Quant Time: Sep 20 03:57:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 20 03:11:07 2024
Response via : Initial Calibration

