

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092018\
 Data File : VV007652.D
 Acq On : 20 Sep 2018 13:36
 Operator : SY/MD
 Sample : J4900-02
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A3ZE3

Quant Time: Sep 21 02:01:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 21 01:37:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	103323	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	93246	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	50276	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	19639	4.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.60%
7) Chloroethane-d5	1.58	69	21126	5.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.00%
11) 1,1-Dichloroethene-d2	2.13	63	38523	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	3.96	46	53737	49.60	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.20%
24) Chloroform-d	4.41	84	49063	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
26) 1,2-Dichloroethane-d4	5.09	65	26480	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.80%
32) Benzene-d6	5.10	84	95835	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.60%
36) 1,2-Dichloropropane-d6	6.12	67	28337	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.36	98	97260	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
43) trans-1,3-Dichloropropene-	7.67	79	10519	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.14	63	40210	47.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.12%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	22912	5.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	43410	5.23	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092018\
Data File : VV007652.D
Acq On : 20 Sep 2018 13:36
Operator : SY/MD
Sample : J4900-02
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A3ZE3

Quant Time: Sep 21 02:01:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091918WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 21 01:37:43 2018
Response via : Initial Calibration

