

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092018\  
 Data File : VV007657.D  
 Acq On : 20 Sep 2018 15:53  
 Operator : SY/MD  
 Sample : J4900-15  
 Misc : 25.0 mL/MSVOA V/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 A3ZD9

Quant Time: Sep 21 02:27:37 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR091918WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Sep 21 01:37:43 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.67  | 114  | 106776   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.90  | 117  | 97439    | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.30 | 152  | 52276    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |        |
|--------------------------------|--------|-------|----------|----------|------|--------|
| 4) Vinyl Chloride-d3           | 1.32   | 65    | 19417    | 4.43     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 88.60% |
| 7) Chloroethane-d5             | 1.58   | 69    | 20271    | 4.92     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 98.40% |
| 11) 1,1-Dichloroethene-d2      | 2.13   | 63    | 36688    | 3.15     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 63.00% |
| 20) 2-Butanone-d5              | 3.96   | 46    | 53434    | 47.73    | ug/L | -0.01  |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 95.46% |
| 24) Chloroform-d               | 4.41   | 84    | 48442    | 4.63     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 92.60% |
| 26) 1,2-Dichloroethane-d4      | 5.09   | 65    | 25304    | 4.89     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 97.80% |
| 32) Benzene-d6                 | 5.10   | 84    | 92822    | 4.81     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 96.20% |
| 36) 1,2-Dichloropropane-d6     | 6.12   | 67    | 27614    | 4.85     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 97.00% |
| 41) Toluene-d8                 | 7.36   | 98    | 94559    | 4.75     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 95.00% |
| 43) trans-1,3-Dichloropropene- | 7.67   | 79    | 10215    | 4.46     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 89.20% |
| 46) 2-Hexanone-d5              | 8.14   | 63    | 40215    | 45.04    | ug/L | 0.00   |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 90.08% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.27  | 84    | 21931    | 4.64     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 92.80% |
| 64) 1,2-Dichlorobenzene-d4     | 11.68  | 152   | 42504    | 4.93     | ug/L | 0.00   |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 98.60% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092018\  
Data File : VV007657.D  
Acq On : 20 Sep 2018 15:53  
Operator : SY/MD  
Sample : J4900-15  
Misc : 25.0 mL/MSVOA V/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
A3ZD9

Quant Time: Sep 21 02:27:37 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR091918WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Fri Sep 21 01:37:43 2018  
Response via : Initial Calibration

